

## EU MDR Declaration of Conformity

<b>Manufacturer</b>	<b>NDD Medizintechnik AG</b> Technoparkstrasse 1 CH-8005 Zürich, Switzerland
<b>Product Family</b>	Spirometers
<b>Basic UDI-DI</b>	764014219SensorKitPX
<b>Single Registration Number</b>	CH-MF-000015550
<b>Authorized Representative</b>	<b>NDD Medizintechnik GmbH</b> Endersbacher Strasse 49 71334 Waiblingen Germany  SRN: DE-AR-000032322
<b>Product Trade Name &amp; Catalogue Number</b>	Spiro-SP TrueFlow Sensor, REF: 2700-1SP
<b>CND code</b>	Z121501 - Spirometry Instruments
<b>Classification</b>	Class IIa according to (EU) 2017/745, Annex VIII, Rule 10
<b>Common Specifications</b>	N/A

We hereby declare our sole responsibility for the EU Declaration of Conformity.

The devices covered by this declaration are in conformity with the European Medical Device Regulation (EU) 2017/745 as well as other relevant Union legislations that make provisions for the issuing of a declaration of conformity.

NDD Medizintechnik AG follows the Conformity Assessment procedure based on a Quality Management pursuant to Regulation (EU) 2017/745, Annex IX, which involves the intervention of the Notified Body:

**TÜV SÜD Product Service GmbH, Notified Body 0123**  
Ridlerstrasse 65, 80339 Munich, Germany

This Declaration of Conformity is applicable to all products listed and released after the Date of Issuance of this Declaration of Conformity, and is valid until the expiry date of the EC Certificate G10 005204 0004 Rev. 01.

  
Andreas Senn,  
Director of Quality Services

  
Michael Bencak,  
CEO

Zurich, 14. May 2024